

**Table 2 Accuracy of the results of susceptibility testing of 146 laboratories sent *Citrobacter freundii***

<b>Antibiotic</b>	<b>Expected</b>	<b>Correct No. of labs</b>	<b>Incorrect No. of labs</b>	<b>No answer No. of labs</b>
Ampicillin	R/I	141	4	1
Cephalexin	R	113	26	7
Cefotaxime	S	95	20	31
Chloramphenicol	S	108	27	11
Gentamicin	S	123	13	10
<b>Total</b>		<b>580 (79.5%)</b>	<b>90 (12.3%)</b>	<b>60 (8.2%)</b>

*R = resistant, S = susceptible, I = intermediate.*